

รหัสโครงการ : RDG5550071
 ชื่อโครงการ : การเตรียมเทอร์โมพลาสติกอิเล็กโทรเมอร์และเทอร์โมพลาสติกวัลค้านีซจากการเบلنด์อะคริลิกเรซินผสานยางธรรมชาติอิพอกซี่ไดซ์เพื่อใช้เป็นวัสดุในการเตรียมฐานพื้นปлом
 ชื่อนักวิจัย : อนุวัตร วงศ์
 สังกัด : สาขาวิชาเคมีประยุกต์ คณะวิทยาศาสตร์และเทคโนโลยีฯ มหาวิทยาลัยอิสลาม ยะลา
 โทรศัพท์ : 073-418609
 E-mail : Anuwatworlee@yahoo.com
 ระยะเวลาโครงการ : 1 สิงหาคม 2555 ถึง 31 กรกฎาคม 2556

บทคัดย่อ

อะคริลิกเรซินจัดได้ว่าเป็นพอลิเมอร์ชนิดหนึ่งที่นิยมนำมาใช้เป็นวัสดุพอลิเมอร์สำหรับใช้แปรรูปฐานพื้นปломมานานหลายปี แต่ก็ยังพบว่าวัสดุดังกล่าวอย่างมีข้อด้อยบางประการที่ส่งผลให้ฐานพื้นปломเกิดการแตกหักได้่ายหลังการใช้งาน เช่น มีความทนทานต่iore แรงกระแทก และแรงตัดโค้งหรือโค้งอต้า ซึ่งเป็นสาเหตุหนึ่งที่ทำให้ฐานพื้นปломที่แปรรูปจากอะคริลิกเรซินเกิดการแตกหักหลังการใช้งานได้่าย ดังนั้นในงานวิจัยนี้จึงจะทำการศึกษาปรับปรุงสมบัติข้อด้อยของฐานพื้นปломอะคริลิกเรซินข้างต้น โดยการเตรียมพอลิเมอร์เบلنด์จากการผสมระหว่างอะคริลิกเรซินและยางธรรมชาติอิพอกซี่ไดซ์ จากการศึกษาพบว่าสมบัติของพอลิเมอร์เบلنด์ที่ปราภูณในกราฟ stress – strain จะมีความแข็งดึงลดลงตามการเพิ่มปริมาณยางธรรมชาติอิพอกซี่ไดซ์ ส่วนสมบัติค่าโมดูลัสตัวโค้งและความทนทานต่iore แรงกระแทกของพอลิเมอร์เบلنด์จะมีค่าสูงกว่าอะคริลิกเรซินบริสุทธิ์เมื่อทำการเบلنด์อะคริลิกเรซินกับยางธรรมชาติอิพอกซี่ไดซ์ในรูปน้ำยา yang คอมแพตในปริมาณร้อยละ 5 โดยน้ำหนักเนื้อยางแห้ง ส่วนการผสมอะคริลิกเรซินกับยางธรรมชาติอิพอกซี่ไดซ์ในอัตราส่วนอื่นๆ พบว่ามีผลทำให้ค่าโมดูลัสตัวโค้งและความทนทานแรงกระแทกลดลง นอกจากนี้ยังพบว่าการผสมเส้นใยไฟในพอลิเมอร์เบلنด์ในปริมาณร้อยละ 10 โดยน้ำหนัก จะทำให้ขึ้นทดสอบพอลิเมอร์เบلنด์มีค่าโมดูลัสตัวโค้งและความทนทานต่iore แรงกระแทก (Impact Strength) เพิ่มขึ้นเมื่อเปรียบเทียบกับขึ้นทดสอบอะคริลิกเรซินบริสุทธิ์

คำสำคัญ: ยางธรรมชาติ ยางธรรมชาติอิพอกซี่ไดซ์ อะคริลิกเรซิน ฐานพื้นปлом

Project code: RDG5550071
Project title: Preparation of Thermoplastic Elastomer and Thermoplastic Vulcanized from Acrylic resin blends with Epoxidized Natural Rubber used to prepared Denture bases

Investigator : Anuwat Worlee
Telephone number: 073-418609
E-mail : Anuwatworlee@yahoo.com
Project duration : 1 August 2012 to 31 July 2013

Abstract

Although acrylic resin, polymethyl methacrylate (PMMA) has been widely used as a main component of denture base polymer for many years, this material is sometime fracture or cracked in clinical use. One of the factors that causes fracture is considered to be low resistance to impact, flexural or fatigue. In this studies, improvement of that inferior properties of denture base with blending acrylic resin and epoxidized natural rubber was investigated. The stress – strain curve shown lower stiffness of polymer blends, acrylic resin and epoxidized natural rubber when increased the quantities of epoxidized natural rubber phase. This is an indication that the elastic properties increased with increasing the quantities of epoxidized natural rubber. The polymer blend shown higher the flexural modulus and impact strength than pure acrylic resin, when blends acrylic resin with 5% by weight of dry rubber content of epoxidized natural rubber latex compound., this properties are lower with had blend in other ratios. In Addition, we can also seen that, the addition of bamboo fiber at 10% wt. in polymer blends cause to the specimen of its gave higher the flexural modulus value and impact strength than pure acrylic resin specimen.

Keyword: Natural Rubber, Epoxideze Natural Rubber, Acrylic Resin, Denture base